

JEWELLED
BEARING

**INSPECTION
CERTIFICATE**
TRACEABLE TO NIST

PRECISION DIAL INDICATORS

- Graduation: 0.001mm
- Jeweled bearing
- Supplied with limit pointers and bezel clamp
- Optional accessory: backs (page 167~168), contact points (page 164~166)



2313-1A

(mm)

Code	Range	Accuracy	Hysteresis	Range/rev	Dial reading	Remark	A	B	C	D
2313-1FA *	1mm	5 μ m	2 μ m	0.2mm	0-100-0	flat back	45	62.5	19.5	15
2313-2FA *	2mm	6 μ m	2.5 μ m	0.2mm	0-100-0	flat back	45	62.5	19.5	15
2313-1A *	1mm	5 μ m	2 μ m	0.2mm	0-100-0	lug back	45	62.5	19.5	15
2313-2A *	2mm	6 μ m	2.5 μ m	0.2mm	0-100-0	lug back	45	62.5	19.5	15

* Supplied with manufacturer inspection certificate traceable to NIST USA

