

JEWELLED
BEARING

**INSPECTION
CERTIFICATE**
TRACEABLE TO NIST

- Meet DIN878
- Suitable for depth and step measurements
- Jeweled bearing
- Supplied with limit pointers and bezel clamp
- Optional accessory: backs (page 167~168), contact points (page 164~166)



2801-10

REVERSE READING DIAL INDICATORS

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading	Remark
2801-10F *	10mm	0.01mm	17 μ m	3 μ m	1mm	100-0	flat back
2801-10 *	10mm	0.01mm	17 μ m	3 μ m	1mm	100-0	lug back

* Supplied with manufacturer inspection certificate traceable to NIST USA

